

The	SCAM4	K Series	Camera Help Manual	1
Con	tents			I
1	SCAM4	4K Serie	s Camera Application	1
2	SCAM4	4K Serie	s Camera Datasheet and Functions (2)	1
3	Dimen	sion of S	SCAM4K Series Camera	3
4	SCAM4	4K Serie	s Camera Packing Information	3
5	Softwa	re and <i>i</i>	Арр	4
6	SCAM4	4K Serie	s Camera Configurations	5
	6.1	Camera	a working standalone with built-in XCamView software	5
	6.2	Connec	cting camera to computers with USB3.0 Port	6
	6.3	Camera	a working in WiFi mode (AP mode)	7
	6.4	Conneo	cting multi-cameras to the router through the WiFi STA mode for the network application	8
7	Brief Ir	ntroduct	ion of SCAM4K UI and Its Functions	11
	7.1	XCamV	/iew UI	11
	7.2	The ca	mera control panel on the left side of the video window	11
	7.3	The Me	easurement Toolbar on top of the video window	11
	7.4	Icons a	nd functions of the Synthesis Camera Control Toolbar at the bottom of the video window	13
	7.	4.1	Setting>Network>General	13
	7.	.4.2	Setting>Network>WiFi	13
	7.	4.3	Setting>Measurement	14
	7.	4.4	Setting>Magnification	14
	7.	4.5	Settings>Image Format	14
	7.	4.6	Setting>Video	15
	7.	4.7	Setting>Storage	15
	7.	.4.8	Setting>Files	15
	7.	4.9	Setting>Time	16
	7.	4.10	Setting>Language	16
	7.	4.11	Setting>ISP	17
	7.	4.12	Setting>Miscellaneous	17
8	Sample	e Photos	s Captured with SCAM4K Series Camera	18
9	Contac	ting Cus	stomer Service	19

Contents

1 SCAM4K Series Camera Application



Figure 1-1 The SCAM4K Series Camera

The SCAM4K series camera is intended for acquisition of digital images from stereo microscopes, biological microscopes, or online interactive teaching. The basic characteristic is listed as below:

- Sony Exmor/STARVIS back-illuminated CMOS sensor
- 4K HDMI/ WiFi/ USB3.0 multiple video outputs
- 4K/1080P auto switching according to monitor resolution
- SD card/USB flash drive for captured image and video storage, support local preview and playback
- Embedded XCamView for the control of the camera and image processing
- Excellent ISP with local tone mapping and 3D denoising
- ToupView/ToupLite software for PC
- iOS/Android applications for smart phones or tablets

2 SCAM4K Series Camera Datasheet and Functions (2)

Order Code	Sensor & Size(mm)	Pixel(µm)	G Sensitivity/Dark Signal	FPS/Resolution	Binning	Exposure(ms)
SCAM4K8MPA	Sony IMX678(C) 1/1.8"(7.68x4.32)	2.0x2.0	1364mv with 1/30s 0.15mv with 1/30s	30@3840*2160(HDMI) 30@1920*1080(WiFi) 30@3840*2160(USB3.0)	1x1	0.045~1000
SCAM4K8MPB	Sony IMX585(C) 1/1.2"(11.14x6.26)	2.9x2.9	1028mv with 1/30s 0.13mv with 1/30s	30@3840*2160(HDMI) 30@1920*1080(WiFi) 30@3840*2160(USB3.0)	1x1	0.014~1000



Figure 2-1 Available Ports on the Back Panel of the Camera Body

Interface or Button	Function Description			
USB Mouse	Connect USB mouse for easy operation with embedded XCamView software			
USB2.0	Connect USB flash drive to save pictures and videos Connect 5G WiFi module to transfer video wirelessly in real time			
USB Video	Connect PC or other host device to realize video image transmission			

	Comply with HDMI1.4 standard. 4K/1080P format video output and supporting automatic switch between 4K and				
HDMI	1080P format according to the connected monitors				
SD	Comply with SDIO3.0 standard and SD card could be inserted for video and images saving				
ON/OFF	Power switch				
LED	LED status indicator				
DC12V	Power adapter connection (12V/1A)				
Video Output Interface	Function Description				
HDMI Interface	Comply with HDMI1.4 standard 30fps@4K or 30fps@1080P				
WiFi Interface	Connecting 5G WiFi adapter (USB2.0 slot) in AP/STA mode 1080P H264 format video, 8M (3840*2160) image				
USB Video Interface	Connecting USB3.0 Video port of PC for video transfer H264/NV12/MJPEG format video				
Other Function	Function Description				
Video Saving	Video format: 8M(3840*2160) H264/H265 encoded MP4 file				
video Saving	Video saving frame rate: 30fps				
Image Capture	8M (3840*2160) JPEG/TIFF image in SD card or USB flash drive				
Measurement Saving	Measurement information saved in different layer with image content				
0	Measurement information is saved together with image content in burn in mode Exposure(Automatic / Manual Exposure) / Gain, White Balance(Manual / Automatic / ROI Mode),				
ISP	Sharpening, 3D Denoise, Saturation Adjustment, Contrast Adjustment, Brightness Adjustment, Gamma				
ISP					
	Adjustment, Color to Gray, 50HZ/60HZ Anti-flicker Function				
Image Operation Zoom In/Zoom Out(Up to 10X), Mirror/Flip, Freeze, Cross Line, Compare(Comparison between real time v and images in SD card or USB flash drive), Embedded Files Browser, Video Playback, Measurement Function					
Embedded RTC(Optional)	To support accurate time on board				
Restore Factory Settings	Restore camera parameters to its factory status				
Multiple Language	guage English / Simplified Chinese / Traditional Chinese / Korean / Thailand / French / German / Japanese / Italian / Russian				
Support					
	Software Environment under WiFi/USB Video Output				
White Balance	Auto White Balance				
Color Technique	Ultra-Fine Color Engine				
Capture/Control SDK	Windows/Linux/macOS/Android Multiple Platform SDK(Native C/C++, C#/VB.NET, Python, Java, DirectShow, Twain, etc)				
Recording System	Still Picture or Movie				
Operating System	Microsoft® Windows® XP / Vista / 7 / 8 / 8.1 /10(32 & 64 bit) OSx(Mac OS X) Linux				
	CPU: Equal to Intel Core2 2.8GHz or Higher				
	Memory: 4GB or More				
PC Requirements	Display:19" or Larger				
	CD-ROM				
	Operating Environment				
Operating Temperature (in Centidegree)	-10°~ 50°				
	-20°~ 60°				
Storage Temperature (in Centidegree)					
Storage Temperature (in Centidegree) Operating Humidity	30~80%RH				
Centidegree)	30~80%RH 10~60%RH				


3 Dimension of SCAM4K Series Camera



Figure 3-1 Dimension of SCAM4K Series Camera

4 SCAM4K Series Camera Packing Information





	Standard Packing List				
Α	Gift box : L:25.5cm W:17.0cm H:9.0cm (1pcs, 1.57Kg/ box)				
В	B SCAM4K Camera(One of the two different shapes)				
С	Power Adapter: Input: AC 100~240V 50Hz/60Hz, Output: DC 12V 1A American standard: Model: POWER-U-12V1A(MSA-C1000IC12.0-12W-US): UL/CE/FCC European standard: Model: POWER-E-12V1A(MSA-C10001C12.0-12W-DE): UL/CE/FCC EMI standard: FCC Part 15 Subpart B EMS standard: EN61000-4-2,3,4,5,6				
D	USB Mouse				
Е	HDMI Cable				
F	USB3.0 A male to A male gold-plated connectors cable /2.0m				
G	CD (Driver & utilities software, Ø12cm)				
	Optional Accessory				
Н	SD Card(16G or above; Speed: class 10)				

I	Adjustable lens adapter C-Mount to Dia.23.2mm Eyepiece Tube (Please choose 1 of them for your microscope		108001/AMA037 108002/AMA050 108003/AMA075			
J	Fixed lens adapter	C-Mount to Dia.23.2mm Eyepiece Tube (Please choose 1 of them for your microscope)	tope) 108005/FMA037 108006/FMA050 108007/FMA075			
	Note: For K and L optional items, please specify your camera type(C-mount, microscope camera or telescope camera), ToupTek engineer will help you to determine the right microscope or telescope camera adapter for your application;					
K	108015(Dia.23.2mm to 30.0mm Ring)/Adapter rings for 30mm eyepiece tube					
L	108016(Dia.23.2mm to 30.5mm Ring)/ Adapter rings for 30.5mm eyepiece tube					
М	Calibration kit		106011/TS-M1(X=0.01mm/100Div.); 106012/TS-M2(X,Y=0.01mm/100Div.); 106013/TS-M7(X=0.01mm/100Div., 0.10mm/100Div.)			
Ν	USB flash drive					
0	USB WiFi adapter					

5 Software and App

The software or the APP can be downloaded from the following link:

Windows: http://www.touptek.com/download/showdownload.php?lang=en&id=33

Linux & macOS: http://www.touptek.com/download/showdownload.php?lang=en&id=28

iOS: <u>https://itunes.apple.com/us/app/toupview/id911644970</u>

Android: https://play.google.com/store/apps/details?id=com.touptek.tpview

6 SCAM4K Series Camera Configurations

You can use the SCAM4K series camera in 4 different ways. Each application requires different hardware environment.


6.1 Camera working standalone with built-in XCamView software

For this application, apart from the microscope, you only need an HDMI monitor, the supplied USB mouse, and the camera embedded XCamView software. A computer or a network connection is not required to operate the camera in this application. The steps to start the camera are listed as below:




Figure 6-1 SCAM4K Series Camera with the HDMI Monitor


Connect the camera to a HDMI monitor using the HDMI cable;



Insert the supplied USB mouse to the camera's USB port;




Insert the supplied SD card/USB flash drive (USB2.0 slot) into the SCAM4K series camera SD card slot/USB2.0 slot;



1 as /

Connect the camera to the power adapter and turn it on;



Turn on the monitor and view the video in the XCamView software. Move the mouse to the left, top or bottom of the XCamView UI, different control panel or toolbar will pop up and users could operate with the mouse at ease.




Figure 6-2 XCamView And SCAM4K Series Camera in HDMI Mode

6.2 Connecting camera to computers with USB3.0 Port

For Windows user (Windows XP (32bit), Windows 7/8/10/11 (32/64 bit)), please use ToupView.

For macOS and Linux user (macOS 10.10 or above or Linux distributions with kernel 2.6.27 or higher), please use ToupLite. The steps to start the camera are listed below:

Start the camera according to Sec. 6.1. After the camera is running, connect camera to computer with USB cable. Please use "USB Video" slot, NOT "USB Mouse" slot as shown below.



Install ToupView/ToupLite on your PC or install ToupView App on the mobile device; Run the software ToupView/ToupLite, clicking the camera name in the camera list group to start the live video as shown in Figure 6-3.




Figure 6-3 ToupView and SCAM4K Series Camera in USB Mode

6.3 Camera working in WiFi mode (AP mode)

Please make sure your PC is WiFi enabled.

For Windows user (Windows XP (32bit), Windows 7/8/10/10/11 (32/64 bit)), please use ToupView.

For macOS and Linux user (macOS 10.10 or above or Linux distributions with kernel 2.6.27 or higher), please use ToupLite. When connecting the camera with a mobile device, the free ToupView App is required. Just make sure that the mobile device uses iOS 11 or higher/Android 5.1 or higher operating systems.


The steps to start the camera are listed below:

Start the camera according to Sec. 6.1. After the camera is running, move the mouse to the bottom of the GUI


and clicking the ²⁰ button on the Synthesis Camera Control Toolbar at the bottom of the video window, a small window called Settings will pop up as shown below. Click Network>WiFi property page and choose the AP in the Wi-Fi Mode edit box(The factory default configuration is AP mode).

×	Settings	
letwork General Wi	Fi	
Intervention (Construction) (Constru	AP 56 *	

Plug the USB WiFi adapter into the camera's USB2 .0 port;



Install ToupView/ToupLite on your PC or install ToupView App on the mobile device, Connect the PC or mobile device to the camera's WiFi AP point; The network name (SSID) and the WiFi password (The default one is 12345678) can be found on the camera's Setting>Network>WiFi page in AP mode.



Start ToupView/ToupLite software or ToupView App and check the configuration. Normally, the active SCAM4K series cameras will be automatically recognized. The live image of each camera is shown in Figure 6-4. For the display, the Camera List tool window is used in ToupView/ToupLite software, and the Camera Thumbnail is used in ToupView App.




Figure 6-4 ToupView and SCAM4K Series Camera in WiFi AP Mode

6.4 Connecting multi-cameras to the router through the WiFi STA mode for the network application

In WiFi STA mode, the camera connects to the router by WiFi STA mode. If a router with WiFi capability is used,


users could connect the router with WiFi to control the camera.

Start the camera according to Sec. 6.1. After the camera is running, move the mouse to the bottom of the video window and clicking the button on the Synthesis Camera Control Toolbar at the bottom of the video window, a small window called Settings will pop up as shown below. Clicking Network>WiFi property page and choosing the STA in the Wi-Fi Mode edit box(The factory default configuration is AP mode). Input the to be connected router's SSID and Password as shown below:


8	Settings
Network	General WiFi
Measurement	WiFi Mode: STA
Magnification	SSID:
Image Format	Password:
Video	
Storage	
Files	
Time	
Language	
ISP	
Miscellaneous	
	Closel Apply

Install ToupView /ToupLite software on your PC. Alternatively, install the free ToupView App on the mobile device;

Plug the USB WiFi adapter into the camera's USB2.0 port;



Finally, as shown below, 4 SCAM4K series cameras are connected to the same router with WiFi STA mode(The number of the cameras, the connection mode(WiFi STA)) connected to the router are determined by the router performance)





Make sure that your PC or your mobile device is connected to the WiFi of the router; Start ToupView/ToupLite software or ToupView App and check the configuration. Normally, active SCAM4K series cameras are automatically recognized. The live image of each camera is displayed. For the display, Camera List control panel window is used in ToupView/ToupLite software, and Camera Thumbnail is used in ToupView App; Select the SCAM4K series camera

you are interested in. To do so, double click the camera's name in Camera List tool window if you use ToupView /ToupLite software; If you use ToupView App, tap the camera's thumbnail in Camera List page(See Figure 6-5)

About the routers/switches

It is suggested that routers/switches supporting 802.11ac 5G segment should be selected to achieve better wireless connection experience.





7 Brief Introduction of SCAM4K UI and Its Functions

7.1 XCamView UI

The SCAM4K UI shown in Figure 6-2 includes a Camera Control Panel on the left of the video window, a Measurement Toolbar on the top of the video window and a Synthesis Camera Control Toolbar on the bottom of the video window.

Notes			
1	To show the Camera Control Panel, move your mouse to the left of the video window. See Sec.7.2 for details		
2	Move the mouse cursor to the top of the video window, a Measurement Toolbar will pop up for calibration and measurement operations. When user left-clicks the Float/Fixed button \checkmark on the Measurement Toolbar, the Measurement Toolbar will be fixed. In this case the Camera Control Panel will not pop up automatically even if users move mouse cursor to left side of the video window. Only when user left-clicks the button on the Measurement Toolbar to exit from measuring procedure will they be able to do other operations on the Camera Control Panel, or the Synthesis Camera Control Toolbar. During the measuring process, when a specific measuring object is selected, an Object Location & Attributes Control Bar $\land \forall \leqslant \geqslant \&$ \textcircled{m} will appear for changing location and properties of the selected object. See Sec.7.3 for details.		
3	When users move mouse cursor to the bottom of the video window, the Synthesis Camera Control Toolbar will pop up automatically. $\oplus \odot \square \bowtie \odot \oplus \odot$		

7.2 The camera control panel on the left side of the video window

The Camera Control Panel controls the camera to achieve the best video or image quality according to the specific applications; It will pop up automatically when the mouse cursor is moved to the left side of the video window (in measurement status, the Camera Control Panel will not pop up. The Camera Control Panel will only pop up when the measurement process is finished or terminated while user's cursor on the left edge of the video window). Left-clicking * button to achieve Display/Auto Hide switch of the Camera Control Panel.

Camera Control Panel	Function	Function Description			
Camera Control Panel	Snap	Capture image and save it to the SD card/ USB flash drive			
Snap Record	Record	Record video and save it to the SD card/ USB flash drive			
Snap Auto Exposure:	Auto Exposure	When Auto Exposure is checked, the system will automatically adjust exposure time and gain according to the value of exposure compensation			
Exposure Compensation: 30	Exposure Compensation	Available when Auto Exposure is checked. Slide to left or right to adjust Exposure Compensation according to the current video brightness to achieve proper brightness value			
Exposure Time: 9ms Gain: 1	Exposure Time	Available when Auto Exposure is unchecked. Slide to left or right to reduce or increase exposure time, adjusting brightness of the video			
White Balance:	Gain	Adjust Gain to reduce or increase brightness of video. The Noise will be reduced or increased accordingly			
Red: 77	Red	Slide to left or right to decrease or increase the proportion of Red in RGB on video			
Green: 31 Blue: 45	Green	Slide to left or right to decrease or increase the proportion of Green in RGB on video			
	Blue	Slide to left or right to decrease or increase the proportion of Blue in RGB on the video			
Sharpness: 32 Denoise: 32	Auto	White Balance adjustment according to the window video every time the button is clicked			
Saturation: 30	Manual	Adjust the Red or Blue item to set the video White Balance			
Gamma: 7 Contrast: 10	ROI	Check the ROI item will display a red ROI rectangle on the video window, drag it to the interested area will perform the White Balance according to the area video data			
Brightness: 50	Sharpness	Adjust Sharpness level of the video			
● DC ○ AC(50Hz) ○ AC(60Hz)	Denoise	Slide left or right to denoise the video			
Default	Saturation	Adjust Saturation level of the video			
Gamma	Adjust Gamma le	vel of the video. Slide to the right side to increase Gamma and to the left to decrease Gamma.			
Contrast	Adjust Contrast level of the video. Slide to the right side to increase Contrast and to the left to decrease Contrast.				
Contrast	Adjust Brightness Brightness.	ightness level of the video. Slide to the right side to increase Brightness and to the left to decrease s.			
DC	For DC illuminati	or DC illumination, there will be no fluctuation in light source so no need for compensating light flickering			
AC(50HZ)	Check AC(50HZ) to eliminate flickering caused by 50Hz illumination				
AC(60HZ)	Check AC(60HZ)	neck AC(60HZ) to eliminate flickering caused by 60Hz illumination			
Default	Restore all the set	the settings in the Camera Control Panel to default values			

7.3 The Measurement Toolbar on top of the video window

The Measurement Toolbar will pop up when moving mouse cursor to any place near the upper edge of the

video window. Here is the introduction of the various functions on the Measurement Toolbar:

≰ ⊠visible Pixel 🛛 MA 🐨 🗶 // · /) − | // × □ ◊ ○ ♡ ⊖ ⊖ ◎ @ & Ø ○ ⊡ ☆ ∫ Ш ↗ 淡 ⊡ ♪ 🛍 ×

Figure 7-1 The Measurement Toolbar on the Upper Side of the Video Window

Icon	Function
La	Float/ Fix switch of the Measurement Toolbar
✓ Visible	Show / Hide Measurement Objects
Pixel ·	Select the desired Measurement Unit
NA ·	Select Magnification for Measurement after Calibration
×	Object Select
K	Angle
\land	4 Points Angle
•	Point
/	Arbitrary Line
$\overline{}$	3 Points Line
/	Horizontal Line
	Vertical Line
X	3 Points Vertical Line
11	Parallel
	Rectangle
\diamond	3 Point Rectangle
0	Ellipse
0 0 0 0 0	5 Point Ellipse
Θ	Circle
0	3 Points Circle
0	Annulus
0	3 Points Annulus
8	Two Circles and its Center Distance
Ø	3 Points Two Circles and its Center Distance
0	Arc
T	Text
	Polygon
S	Curve
um	Scale Bar
$\overline{}$	Arrow
	Execute Calibration to determine the corresponding relation between magnification and resolution, which will establish the corresponding relationship between measurement unit and the sensor pixel size. Calibration needs to be done with the help of a micrometer. For detailed steps of carrying out Calibration please refer to ToupView help manual.
	Export the Measurement information to CSV file(*.csv)
<u> </u>	Measurement Setup
	Delete all the measurement objects
×	Exit from Measurement mode
& ♥ < > ▲ i	When the measurement ends, left-click on a single measuring object and the Object Location & Properties Control Bar will show up. User could move the object by dragging the object with the mouse. But more accurate movement could be done with the control bar. The icons on the control bar mean Move Left, Move Right, Move Up, Move Down, Color Adjustment and Delete.

Note:

1) When user left-clicks Display/Hide button on Measurement Toolbar, Measurement Toolbar will be fixed. In this case Camera Control Panel will not pop up automatically even if moving the mouse cursor to the left edge of the video window. Only when user left-click the button on Measurement Toolbar to exit from the measurement mode will they be able to doing other operations on Camera Control Panel or Synthesis Camera

Control Toolbar.

2) When a specific Measurement Object is selected during the measurement process, Object Location & Attributes Control Bar 🖉 < 🏷 📥 👼 will appear for changing the object location and properties of the selected objects.

Icons and functions of the Synthesis Camera Control Toolbar at the bottom of the video 7.4 window

$\oplus \ominus \bot \ominus \odot \odot \# \odot \odot \exists \& \odot$

Figure 7-2 The Synthesis Camera Control Toolbar on the Bottom of the Video Window

Icon	Function	Icon	Function
\oplus	Zoom In the Video Window	Θ	Zoom Out the Video Window
	Horizontal Flip		Vertical Flip
(C+G)	Color/gray		Video Freeze
#	Display Cross Line		Image Overlay
	Compare Image with the Current Video	67	Browse images and videos in the SD Card
X	Settings	i	Check the Version of XCamView

The 🔀 Setting function is relatively more complicated than the other functions. Here is more information about it:

Setting>Network>General 7.4.1

8	Settings	×
Network	General WiFi	
Measurement		
Magnification	Name: SCAM4K8MPA	
Image Format		
Video		
Storage		
Files		
Time		
Language		
ISP		
Miscellaneous		
		Close Apply

		Figure 7-3 Comprehens	sive Ne	etwork Ge	neral Settings Page	
Name		The current camera name recognized as t	he netwo	ork name		
7.4.2 Sett	ting>l	Network>WiFi				
Wi-Fi Mode		AP/STA mode to select;				
Channel/SSID)	Channel for the AP mode and SSID for the	STA mo	de. Here, the	SSID is the router's SSID;	
Password		Camera Password for the AP mode. Router	Passwor	rd for the STA	A mode	
8		Settings	×	8	Settings	×
Measurement W Magnification F Image Format C	eneral WiF (iFi Mode: [Frequency: 5 Channel: 2 Password: 7	AP • 56 • 36 •		Network Measurement Magnification Image Format Video Storage Files Time Language ISP Miscellaneous	Remeral HIF1 HIF1 SSID: Password:	
		2	Close Apply		Close	Apply

Figure 7-4 Network Setup

7.4.3 Setting>Measurement

This page is used for the define of the Measurement Object properties.

8	Settings	×
Network	Global	1
Measurement	Precision The Calculation results keep 20 decimals Font Size Large	
Magnification Image Format Video	© None Cursor © Single Cross	
Storage Files Time	Hiscellaneo. Wide the label when moving the measurement object © Calibration Line Width 2 Color	
Language ISP Miscellanenus	© Angle Line Width 2 Color Label Tupe ⊠Angle	•
	Line Width 2 Color Label Type ☑ Position ≑Line	÷-
	Line Width 2 Color Label Tune Ølength	Default
		Close Apply

Figure 7-5 The Measurement Setup

Global	Used for setting	digits behind the decimal point for measurement results;
	Line Width	Used for defining width of the lines for calibration;
Calibration	Color	Used for defining color of the lines for calibration;
Cultoration	EndPoint	Type: Used for defining shape of the endpoints of lines for calibration: Null means no EndPoint, rectangle means rectangle type of endpoints. It makes alignment more easily;
Point, Angle, Line	e, Horizontal Line	, Vertical Line, Rectangle, Circle, Ellipse, Annulus, Two Circles, Polygon, Curve
	C	- along with the Measurement command mentioned above will unfold the corresponding attribute settings to I property of the Measurement Objects.

7.4.4 Setting>Magnification

This page's items are formed by the Measurement Toolbar's Calibration command.

Name	Names such as 10X, 40X, 100X are based on magnification of the microscopes. For continuous zoom microscopes, ensure that the selected magnification coincides with the scale alignment line on the microscope zoom knob; Users could also edit the name of the magnification with other information, for example, microscope mode, users name, etc.
Resolution	Pixels per meter. Image device like microscopes have high Resolution value;
Clear All	Click the Clear All button will clear the calibrated magnifications;
Delete	Click Delete to delete the selected magnification;




Figure 7-6 Comprehensive Magnification Settings Page

7.4.5 Settings>Image Format

Image Format	JPEG: The extension of JPEG file can get very high compression rate and display very rich and vivid images by removing redundant images and color data. In other words, it can get better image quality with the least disk space. If measurement objects are available, the measurement objects will be burned into the image and the measurement cannot be edited. TIFF: TIFF is a flexible bitmap format mainly used to store images including photos and artistic images.
Measurement Object Saving Method	Burn in Mode: The measurement objects are merged into the current image. User could not edit the measurement objects any more. This mode is not reversable. Layered Mode: The measurement objects are saved in different layer with current image data in the target file. User could edit the measurement objects in the target file with some software on the PC. This mode is reversable.

8	Settings	×
Network Measurement Magnification Image Format Video Storage Files Time Language ISP Miscellaneous	Settings Image Format © JFE6 O TIFF Measurement Object Saving Method O Burn In Mode @ Layered Mode Burn In Mode Measurement objects are merged into the image. User could not edit the measurement objects any more. Layered Mode Measurement objects are saved in different layer with image data in the target file. User could edit the measurement objects in the target file with software on the PC.	
	Close AD	nly

Figure 7-7 Comprehensive Image Format Settings Page

7.4.6 Setting>Video




Figure 7-8 Comprehensive Setting of Video page

7.4.7 Setting>Storage




Figure 7-9 Comprehensive Setting of Storage Page

File System Format of the Storage Device	List the file system format of the current storage device FAT32: The file system of SD Card is FAT32. The maximum video file size of single file in FAT32 file system is 4G Bytes; exFAT: The file system of SD Card is exFAT. The maximum video file size of single file in FAT32 file system is 16E Bytes; NTFS: The file system of SD Card is NTFS. The maximum video file size of single file is 2T Bytes.
	Unknown Status: SD Card not detected or the file system is not identified;
Note: For USB Flas	sh Drive, USB 3.0 interface is preferred.

7.4.8 Setting>Files

8	Si	ettings	×
Image: State of the second	Si Image File Name @ Auto Prefix : IMG Video File Name @ Auto Prefix : VID	⊘Manual ⊘Manual	×
			Close Apply

Figure 7-10 Comprehensive Setting of Files Name

Image or Video File Name Paradigm	Provide Auto or Manual naming paradigm for Image or Video file;
Auto	With specified name as the Prefix and XCamView will add digital after the Prefix for the Image or Video file;
Manual	A file dialog will pop up to enter the Image or Video file name for the captured Image or Video.

7.4.9 Setting>Time

8		Settings	×
Network	2022-09	-30 15:11:59	
Measurement	Year:	2022	
Magnification	Month:	9 :	
Image Format	Day:	30	
Video	Hour:	15 .	
Storage	Minute:		
Files	Second:	58 -	
Time			
Language			
ISP			
Miscellaneous			
		Clos	se Apply

Figure 7-11 Time Setting

User can set Year, Month, Day, Hour, Minute and Second ital.in this page.

7.4.10 Setting>Language

Time

Network e English e Engl

Figure 7-12 Comprehensive Setting of Language Selection Setting Page

English	Set language of the whole software into English;
Simplified Chinese	Set language of the whole software into Simplified Chinese;
Traditional Chinese	Set language of the whole software into Traditional Chinese;
Korean:	Set language of the whole software into Korean;
Thailand	Set language of the whole software into Thailand;
French	Set language of the whole software into French;

German	Set language of the whole software into German;
Japanese	Set language of the whole software into Japanese;
Italian	Set language of the whole software into Italian;
Russian	Set language of the whole software into Russian;
Dutch	Set language of the whole software into Dutch;
Portuguese	Set language of the whole software into Portuguese;

7.4.11 Setting>ISP

8	Settings	×
Network Measurement Magnification Image Format Video Storage Files Time	Dark Enhance Dark Enhance: ∅ ∄ Auto Exposure Maximum exposure time: 33 ∄ ms ROI Color: □ Synchronized display as Camera Control Panel	
Language ISP Miscellaneous		
	01	ose Apply

Figure 7-13 Comprehensive Setting of ISP Page

Dark Enhance	Define the intensity value of dark enhancement;
Auto Exposure	Define the maximum automatic exposure time;
ROI Color	Choosing the ROI rectangle line color

7.4.12 Setting>Miscellaneous

8	Settings	×
Network	Ruler	
Measurement	□ Show	
Magnification	Measurement	
Image Format	⊠Enable	
Video	Overlay	
Storage	□ Support saving overlay information in Burn In	Mode
Files	Grids	
Time	□ Support saving grids information in Burn In M	nde
Language		
ISP	Cursor	
Miscellaneous	Size: Middle •	
	Camera Parameters	
	Import	Export
	Reset to factory defau	ults
		Close Apply

Figure 7-14 Comprehensive Miscellaneous Settings Page

Ruler	Select to display the ruler in the video window, otherwise not to display the ruler;	
Measurement	Select to display the measurement toolbar in the video window, otherwise not to display the measurement toolbar;	
Overlay	Select to support saving graphics overlay information in fusion mode, otherwise it will not support;	
Grids	Select to support saving mesh information in fusion mode, otherwise not to support;	
Cursor	Choosing the Cursor size according to the screen resolution or personal preference	
Camera Parameters Import	Import the Camera Parameters from the SD Card or USB flash drive to use the previously exported Camera Parameters	
Camera Parameters Export	Export the Camera Parameters to the SD Card or USB flash drive to use the previously exported Camera Parameters	
Reset to factory defaults	Restore camera parameters to its factory status;	

8 Sample Photos Captured with SCAM4K Series Camera




Figure 8-1 Longitudinal Section Of Equisetum Sporophyll Captured with SCAM4K8MPA



Figure 8-2 Lime Wood Stem CS Captured with SCAM4K8MPA




Figure 8-3 Taste Bud.Sec. Captured with SCAM4K8MPA




Figure 8-4 Circuit Board Captured with SCAM4K8MPA

9 Contacting Customer Service

Please contact your local distributor if you have any questions about the product.